

# Devices on FHIR PoCD IG comment list in numerical order

See also [Categorized list of Devices on FHIR PoCD IG comments](#)

- [15334](#) Add use cases showing move from IEEE to FHIR - 2018-Jan PoCD #1 ()
- [15335](#) Explain why PoCD uses Device and PHD doesn't - 2018-Jan PoCD #2 ()
- [15336](#) Why does PoCD use Sample Array while PHD uses Real-Time Sample Array - 2018-Jan PoCD #3 ()
- [15337](#) Why does PoCD use Sample Array while PHD uses Real-Time Sample Array - 2018-Jan PoCD #4 ()
- [15338](#) Reword Rosetta Terminology Mapping Management System explanation - 2018-Jan PoCD #5 ()
- [15339](#) Typo - 2018-Jan PoCD #6 ()
- [15340](#) Use Cases Should Support Full use of the Profiled Resources - 2018-Jan PoCD #7 ()
- [15341](#) Remove DeviceComponent.languageCode in favor of Resource.language - 2018-Jan PoCD #8 ()
- [15342](#) Change DeviceMetric.color & DeviceComponent Channel profile color extension to have a data type of CodeableConcept - 2018-Jan PoCD #9 ()
- [15343](#) Loosen selected bindings to a value set vs. a Fixed Value - 2018-Jan PoCD #10 ()
- [15344](#) Add 11073MDC page to the FHIR standard covering the ISO/IEEE 11073 terminology. - 2018-Jan PoCD #11 ()
- [15345](#) Why does PoCD use Sample Array while PHD uses Real-Time Sample Array - 2018-Jan PoCD #12 ()
- [15346](#) Add to status field values: "paused", "online", "offline" - 2018-Jan PoCD #13 ()
- [15347](#) How is a "gateway" identified - 2018-Jan PoCD #14 ()
- [15348](#) If system both storing and exposing Device data, should there be another "Source"? - 2018-Jan PoCD #15 ()
- [15349](#) operationalStatus should be modeled as IEEE 11073 concept - 2018-Jan PoCD #16 ()
- [15350](#) DeviceComponent languageCode field not needed - 2018-Jan PoCD #17 ()
- [15351](#) Field "color" should use more appropriate coding scheme - 2018-Jan PoCD #18 ()
- [15352](#) Field "identifier" should be conditional or removed - 2018-Jan PoCD #19 ()
- [15353](#) Add dataType field - 2018-Jan PoCD #20 ()
- [15354](#) Text fields support should be explained - 2018-Jan PoCD #21 ()
- [15355](#) Additional use cases suggested - 2018-Jan PoCD #22 ()
- [15356](#) Short background to DoF project needed - 2018-Jan PoCD #23 ()
- [15357](#) Add Scope description - 2018-Jan PoCD #24 ()
- [15358](#) Reference UDI DAM and UDI Pattern ID - 2018-Jan PoCD #25 ()
- [15359](#) Reference UDI DAM and UDI Pattern ID in future documents - 2018-Jan PoCD #26 ()
- [15360](#) show differential view - 2018-Jan PoCD #27 ()
- [15361](#) Un-crowd menu bar - 2018-Jan PoCD #28 ()
- [15362](#) Improve dropdown items and give jumps - 2018-Jan PoCD #29 ()
- [15363](#) Add diagrams to use cases - 2018-Jan PoCD #30 ()
- [15364](#) type slices not rendering correctly in tree view. - 2018-Jan PoCD #31 ()
- [15365](#) Add CapabilityStatement - 2018-Jan PoCD #32 ()
- [15366](#) 33 Datatypes do not belong in terminology - 2018-Jan PoCD #33 ()
- [15367](#) Missing ValueSets referenced - 2018-Jan PoCD #34 ()
- [15368](#) IG format made it difficult to review - 2018-Jan PoCD #35 ()
- [15369](#) Reword Overview text - 2018-Jan PoCD #36 ()
- [15370](#) typo - 2018-Jan PoCD #37 ()
- [15371](#) Terminology resources incomplete - 2018-Jan PoCD #38 ()
- [15372](#) Explain relationship between this IG and the PHD IG - 2018-Jan PoCD #39 ()
- [15373](#) Clarify requirement (or not) for MDC code - 2018-Jan PoCD #40 ()
- [15374](#) Measurement status should be an extension on DeviceMetric - 2018-Jan PoCD #41 ()
- [15375](#) Measurement status should be modifier extension in DeviceMetric - 2018-Jan PoCD #42 ()
- [15376](#) Consolidate introductory sections on Purpose, Devices on FHIR initiative, and Audience - 2018-Jan PoCD #43 ()
- [15377](#) Change stylesheet for IG (created on Forge) to match IG Publisher - 2018-Jan PoCD #44 ()
- [15378](#) Consider "must support" flag on items such as the identifiers - 2018-Jan PoCD #45 ()