

# 2018-07-18 DoF Weekly Meeting

## Date

18 Jul 2018

## Attendees

- [John Rhoads](#) Philips
- Chris Courville - Epic
- Stefan Karl - Philips
- Michael Faughn - Prometheus / NIST
- Koichiro Matsumoto - Nihon Kohden
- Uwe Meyer (M2IT)
- Brian Reinhold (LNI / Continua)
- John Dyer (Draeger)
- Joe Quinn (Optum)
- Ana Kostadinovska (Philips)

## Goals

- Discuss revised Device spreadsheet from [OO Conf Call Minutes - Devices - 20180717](#)  
Reference spreadsheet contained in those Minutes

## Discussion items

Time	Item	Who	Notes
20min	Does the completed part of the spreadsheet (excluding specialization) meet known HCD needs?	John, Brian, Stefan and others	<ul style="list-style-type: none"><li>• safety characteristics discussion - needed or wanted in device instance data? Brian: not needed for PHD, Stefan and John - need not foreseen for PoCD. Probably most useful as a DeviceDefinition element because not it typically does not vary between instances of a kind in HCD use cases, but some other use cases could need it. No objection.</li><li>• patient discussion: comment saying "Patient affixed to" led some to question if it is applicable to communicating devices. Stefan: But patient reference very important for Point-of-Care Devices as patient for whom observations are being generated. Maybe description should make this clear?</li><li>• After discussion, the group was asked if they thought that the spreadsheet as discussed so far was satisfactory for HCD intended uses. All agreed that it was.</li></ul>
	Specialization field: what should be communicated to OO as the consensus of HCD	Brian, Ana	<ul style="list-style-type: none"><li>• Possible workarounds for dispensing with it in representing single-specialization devices. Consensus was that this was not satisfactory and we should communicate to OO that we would like to see this maintained as a field in the core resource rather than dealing with it as an extension.</li><li>• Specialization.version should be changed from cardinality 1..1 to 0..1. Though a version is required in PHD usage, that should be required in a profile rather than in the core resource since there may well be situations where it is not needed or not available.</li></ul>

## Action items

- Communicate results of discussion back to Hans Buitendijk for OOJohn Rhoads